

BD Phoenix™ Automated Microbiology System Comparative System Maintenance



BD Phoenix™ System

DAILY
Check printer paper supply
Record temperature
BD PHOENIX AP DAILY MAINTENANCE
Check consumables - replace if needed
Daily Nephelometer check
WEEKLY
Check LEDs
BD PHOENIX AP WEEKLY MAINTENANCE
Calibrate dispense pump
BD PHOENIX AP MONTHLY MAINTENANCE
Replace dispense tubing set
Calibrate Nephelometer (every 3 months)
AS NEEDED
Pipettor verification as needed
At 6 months (Performed by BD Engineers)
Check/clean or replace air filters
Annually (Performed by BD Engineers)
BD Phoenix AP preventive maintenance

VITEK® 2

DAILY	
Check instrument status QC report for: Carousel temperature A and B Optics system A and B	
NO WEEKLY MAINTENANCE	
MONTHLY	
Empty Vitek 2 Turn power off	
▼ ▼	
Clean the following with 10% bleach solution:	
Boats	
Carousel A and B	
Cassettes	
Optics A and B	
Waste tray A and B	
Base pan	
Vacuum seal	
Vacuum chamber	
Drip pan	
Smart carrier	
AS NEEDED	
Change saline	
▼	
Change pipette tips	

VITEK®

DAILY	
Brinkman Dispenser:	Reader/Incubator:
▼	▼
Clean	Record temperature
▼	
Flush tubing	
▼	
Calibrate	
▼	
Change diluent	
▼	
Waste tray A and B	
WEEKLY	
Brinkman Dispenser:	Reader/Incubator:
▼	▼
Sterilize	Inspect arrays
	▼
	Clean wheels
	▼
	Clean track
MONTHLY	
Filler/Sealer:	Reader/Incubator:
▼	▼
Check timing 3 to 5 minutes	Clean 4 filters
▼	
Clean filling racks	
▼	
Clean door gasket	
NO AS NEEDED MAINTENANCE	

Microscan® WalkAway® SI

DAILY
Print daily instrument QC diagnostics report
▼
Review internal instrument checks
▼
Manually check the following:
Internal and external temperature, water level, reagent bottle level, dispenser pressure, oil level, oil and reagent lines, dispensing tips and funnel, reagent waste level, photodiode shield, diffuser plate
Rapid panel users only: Check fluorometer shield and fluorometer reference disk
NO WEEKLY MAINTENANCE
MONTHLY
Clean fan filter
▼
Clean bottle adapter threads
▼
Clean reagent bottle O rings
AS NEEDED
Print daily instrument QC diagnostics report
▼
Fill water level
▼
Fill or replace reagents
▼
Fill or replace oil
▼
Replace reagent waste bag
▼
Clean fluorometer shield
▼
Clean diffuser plate



BD Diagnostics
7 Loveton Circle
Sparks, MD 21152-0999
800.638.8663
www.bd.com/ds

VITEK is a registered trademark of bioMérieux Inc. MicroScan and WalkAway are registered trademarks of Dade Behring Inc. BD, BD Logo and Phoenix are trademarks of Becton, Dickinson and Company. ©2009 BD. 1-2697 January 2009. Printed in USA