

BD Product Name: BD Pyxis Medstation Anesthesia Station ES

Date of Critical or Security Patches: July 2019

Abstract: Compatablitiy issue with KB405221

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for July 2019.

The patches include updates that may adversely affect BD Pyxis Anesthesia Station ES. These patches will not be and should not be installed on BD Pyxis Anesthesia Station ES at this time.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).

Note: These patches only apply to the latest supported version(s) of the BD offering

Patch Name	Patch Description	Patch ID	Notes
Security update for SQL Server 2016 SP1 CU15	A remote code execution vulnerability exists in Microsoft SQL Server when it incorrectly handles processing of internal functions. An attacker who successfully exploited this vulnerability could execute code in the context of the SQL Server Database Engine service account. To learn more about the vulnerability, go to CVE-2019-1068.	<a href="#">KB4505221</a>	1

Notes

1. Reported on BD Pyxis Medstation ES version 3.22.0.21 and BD Pyxis Anesthesia Station ES version 1.15.0.7. Application crashes on startup after patch is installed. Patch is not approved for installation.

