

Security Patches: BD FACSCanto™ II System

December 2022

BD Biosciences - BD FACSCanto™ II Cell Analyzers

Microsoft™ and Third-Party Patches: October 2022 – December 2022

BD has identified patches from Microsoft™ that have been identified as critical or security related for October 2022 – December 2022. These patches were not found to adversely affect BD products and will be applied according to customers' service agreements.

Customers that maintain patches independent of BD automated delivery should ensure the validated patches are installed on their BD systems as the acting responsible entity in order to maintain the correct security posture of the system(s).

Microsoft™ patches: Windows™ 10 Enterprise LTSC

Patch name	Description	Patch ID	Notes
October 2022 Monthly Cumulative Updates	Security and quality updates for Microsoft™ Windows™ 10 Enterprise (LTSC) v1809	KB5018419	1, 2
November 2022 Monthly Cumulative Updates	Security and quality updates for Microsoft™ Windows™ 10 Enterprise (LTSC) v1809	KB5019966	1, 3
December 2022 Monthly Cumulative Updates	Security and quality updates for Microsoft™ Windows™ 10 Enterprise (LTSC) v1809	KB5021237	1, 4



Microsoft™ patches: Windows™ 10 Professional

Patch name	Description	Patch ID	Notes
October 2022 Monthly Cumulative Updates	Security and quality updates for Microsoft™ Windows™ 10 Professional v21H2	KB5018410	1, 2
November 2022 Monthly Cumulative Updates	Security and quality updates for Microsoft™ Windows™ 10 Professional v22H2	KB5019959	1, 3
December 2022 Monthly Cumulative Updates	Security and quality updates for Microsoft™ Windows™ 10 Professional v22H2	KB5021233	1, 4

Notes

1. Patches may require the system to reboot.
2. For further information from Microsoft regarding CVEs covered by the October Security updates, please refer to the following link – <https://msrc.microsoft.com/update-guide/releaseNote/2022-Oct>
3. For further information from Microsoft regarding CVEs covered by the November Security updates, please refer to the following link – <https://msrc.microsoft.com/update-guide/releaseNote/2022-Nov>
4. For further information from Microsoft regarding CVEs covered by the December Security updates, please refer to the following link – <https://msrc.microsoft.com/update-guide/releaseNote/2022-Dec>

BD FACSCanto™ II Flow Cytometers are Class 1 Laser Products.

In the U.S., the BD FACSCanto™ II Flow Cytometer is for In Vitro Diagnostic Use for up to six colors. Seven and eight colors are for Research Use Only.

In the E.U., the BD FACSCanto™ II Flow Cytometer is no longer available for sale.

BD Biosciences, San Jose, CA, 95131, U.S.
877.232.8995

bd.com

BD, the BD Logo and FACSCanto are trademarks of Becton, Dickinson and Company or its affiliates.
All other trademarks are the property of their respective owners. © 2023 BD. All rights reserved. BD-82673 (v1.0) 0223

