

BD Product Name: BD Pyxis MedStation ES 1.5.2.1. Station for Department of Defense Customers

September 2019

Abstract: Updates not installing on BD Pyxis MedStation ES 1.5.2.1 Station for DoD Customers
Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for September 2019.

The patches include updates that may adversely affect BD Pyxis MedStation ES 1.5.2.1 Station for Department of Defense customers. These patches will not be and should not be installed on BD Pyxis MedStation ES 1.5.2.1 at this time.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).

Note: These patches only apply to the latest supported version(s) of the BD offering

Notes

1. Unable to install updates on this vm environmnet at this time.

