

BD Product Name: **BD Pyxis™ Anesthesia Station ES**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020 will not install on BD Pyxis™ Anesthesia Station ES.

The patch includes updates that may adversely affect Windows 10 Anesthesia Station ES. This patch will not be and should not be installed on BD Pyxis™ Anesthesia Station ES at this time.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).

Note: This patch only apply to the latest supported version(s) of the BD offering.

Patch Name	Description	Patch ID	Notes
2020-11 Cumulative Update for Windows 10 Version 1607 for x64-based Systems	A security issue has been identified in a Microsoft software product that could affect your system. You can help protect your system by installing this update from Microsoft.	KB4586830	None
Windows Malicious Software Removal Tool x64 - v5.84	The Windows Malicious Software Removal Tool (MSRT) helps remove malicious software from computers	KB890830	None



BD Product Name: **BD Pyxis™ Supply**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020 will not install on BD Pyxis™ Supply.

The patch includes updates that may adversely affect Pyxis™ Supply. This patch will not be and should not be installed on BD Pyxis™ Supply at this time.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Alaris™ Systems Manager**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020 will not install on BD Alaris™ Systems Manager.

The patch includes updates that may adversely affect Alaris™ Systems Manager. This patch will not be and should not be installed on BD Alaris™ Systems Manager at this time.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Pyxis™ MedStation™ ES**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches - November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Pyxis™ CII Safe ES**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Care Coordination Engine**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Pyxis™ MedStation™ 4000**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Pyxis™ MedStation™ 3500**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Pyxis™ Anesthesia system 3500**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Pyxis™ Security Module**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Pyxis™ Anesthesia Station 4000**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD HealthSight™ Viewer**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Pyxis™ Order Viewer**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Pyxis™ Logistics**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Pyxis™ Data Agent**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **Identity Management**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure these actions are not performed as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Pyxis™ CIISafe™**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure the validated patches are installed on their BD systems as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Pyxis™ CUBIE™ Replenishment System**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure the validated patches are installed on their BD systems as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Pyxis™ PharmoPack™**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure the validated patches are installed on their BD systems as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Pyxis™ Barcode Labeling**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure the validated patches are installed on their BD systems as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Pyxis™ IV Prep**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure the validated patches are installed on their BD systems as the acting responsible entity in order to maintain the correct security posture of the system(s).



BD Product Name: **BD Health Sight™ Patient Association**

Date of Critical or Security Patches: November 2020

Abstract: Critical or Security Patches – November 2020

Microsoft® & Third-Party Patches

BD has identified patches from Microsoft® that have been identified as critical or security related for November 2020. Patches delivered in the month were not found to adversely affect this BD product and will be applied according to customers' service agreement.

Customers that maintain patches independent of BD automated delivery should ensure the validated patches are installed on their BD systems as the acting responsible entity in order to maintain the correct security posture of the system(s).

