





# BD PosiFlush™ Saline and Heparin Lock Flush Syringes

## Usage guidelines



### Before use:

1. Tear open package at either end and remove the syringe.
2. Depress the plunger with the **tip cap on** to relieve the resistance between the stopper and the barrel (*Fig. a*).
3. Using aseptic technique, remove the syringe tip cap from the **flush** syringe by twisting it off (*Fig. b*).
4. Hold the syringe upright and expel the air and any excess solution in the syringe by positioning the front rib of the gray plunger tip at the desired solution volume (*Fig. c*).
5. Attach the syringe to the injection site and flush, following your institution's policy.

### After use:

6. Discard the used syringe, including any unused solution, following your institution's policy (*Fig. d*). **Do not reuse.**

**Do not** place on a sterile field.  
Please read labeling at **all times**.

**Note:** 3, 5 and 10 mL BD PosiFlush™ syringes all have the same diameter as a standard 10 mL syringe and may be used for flushing catheters when protocols stipulate use of a 10 mL syringe to minimize flush pressure (psi).

BD, Franklin Lakes, NJ, 07417, U.S.

[bd.com](http://bd.com)

BD, the BD Logo and PosiFlush are trademarks of Becton, Dickinson and Company or its affiliates. © 2019 BD. All rights reserved. BD-0932 04/19

